Location:
Department:

JOHNS HOPKINS CIDEX OPA STRIPS WEEKLY QA LOG

NEGATIVE Control is equal to one part Cidex® OPA (15 ml) and one part water (15 ml)
POSITIVE Control is full strength Cidex® OPA (30 ml)
Timing is critical for this test!
Insert strips (3) in each control solution for one (1) seconds
Read results ninety (90) seconds after strips removed from solution.

TEST DATE	SOLUTION LOT NUMBER AND MANU-FACTURER'S EXPIRATION DATE	TEST STRIP LOT NUMBER AND MANU-FACTURER'S EXPIRATION DATE.	CIDEX® OPA STRIPS: DATE OPENED AND EXPIRATION DATE **BASED ON DATE STRIPS OPENED. Strips expire ninety (90) days after container is opened.**	**NEGATIVE** CONTROL RESULT PASS / FAIL **DESIRED RESULT IS FAIL** (Pass = Strip pad is entirely purple; **Fail = Strip pad is blue/ purple.)**	**POSITIVE** CONTROL RESULT PASS / FAIL **DESIRED RESULT IS PASS. (Pass = Strip pad is purple.** If any blue appears on indicating pad apart from the top line, the solution has failed.)	CONTROL PERFORMED BY (INITIALS). Sign name (with initials) once at bottom of log.	**COMMENTS**: **D** = Cidex® OPA Solution Test Strips discarded today because they failed QA test. **D1** = Cidex® OPA Solution Test Strips discarded because cap was left open. **D2** = Cidex® OPA Solution Test Strips discarded because bottle was not dated when it was opened.
9/16/2004	Lot 0-1234 Exp 2004-12	Lot 5-5554 Exp 2006-11	OPENED EXPIRES	FAIL	PASS	LJS	**D1**
Initials:		Signature:					
Initials:		Signature:					
Initials:		Signature:					

THE CORRECT EXPIRATION DATE [MANUFACTURER'S EXPIRATION DATE COMPARED TO DATE BASED ON BOTTLE OF STRIPS BEING OPEN FOR NINETY (90) DAYS] IS ALWAYS THE EXPIRATION DATE WHICH OCCURS FIRST!

Rev. 05/26/06

If you have two bottles of just opened Cidex OPA strips fail the QA test, document your actions on the back of this log.